☐ BIOMETHANE 3000

TECHNICAL SPECIFICATIONS

GENERAL SPECIFICATION	N						
Number of sampling points	1	1					
Gases to be monitored	CH ₄ , CO ₂ and O ₂ wi	CH ₄ , CO ₂ and O ₂ with optional H ₂ S, H ₂ and CO (choice of up to 4)					
Reading intervals	Continuous ¹ CH ₄ , CO ₂ and O ₂ measurement with user definable fourth gas reading						
Operating temperature range	-20°C to +50°C						
POWER							
Mains options	110-230 Vac 50/60 Hz						
Consumption	155W maximum						
Backup memory	Lithium manganes	Lithium manganese dioxide backup battery for memory retention					
GAS RANGES							
Gases measured	CH ₄ and CO ₂	By dual wavelength	nfrared cell with reference channel				
	O ₂	By internal electroch	chemical cell				
	H ₂ S / H ₂ / CO	By external electroc	By external electrochemical cell				
	Cell	Range	Typical accuracy	Typical accuracy (range : accuracy)*		·)*	
Standard gas cells	CH ₄	0-100%	0-100% : ±0.5%	0-100% : ±0.5% (vol)			
	CO ₂	0-100%	0-60% : ±0.5% (\	0-60% : ±0.5% (vol) 60		% : ±1.5% (vol)	
	O ₂	0-25%	0-1%: ±0.05% (vol)			2-25%: ±1.0% (vol)	
	Cell	Range	Typical accuracy (range : accuracy)*				
			Module cell		System cell		
Optional gas cells	H ₂ S	0-50ppm	±1.5% FS	±1.5% FS		±1.5% FS	
	H ₂ S	0-200ppm	±2.0% FS	±2.0% FS		±1.5% FS	
	H ₂ S	0-500ppm	±2.0% FS	±2.0% FS		±2.0% FS	
	H ₂ S	0-1,000ppm	±2.0% FS	±2.0% FS		±2.0%	
	H ₂ S	0-5,000ppm	±2.0% FS	±2.0% FS		±100ppm or 5% of reading (if greater)	
	H ₂ S	0-10,000ppm	±5.0% FS			±200ppm or 5% of reading (if greater)	
	СО	0-1,000ppm	±2.0% FS	±2.0% FS		±3.0% FS	
	H ₂	0-1,000ppm	±2.5% FS	±2.5% FS		±1.5%	
	Range	Response time	Range		Respon	se time	
Response time, T90**	CH ₄	≤10 seconds	H ₂ S (0-50ppm)	H ₂ S (0-50ppm)		≤30 seconds	
	CO ₂	≤10 seconds	H ₂ S (0-200ppm)	H ₂ S (0-200ppm)		≤35 seconds	
	O ₂	≤10 seconds	H ₂ S (0-500ppm)	H ₂ S (0-500ppm) <3		≤35 seconds	
	H ₂	<90 seconds	H ₂ S (0-1,000ppm	H ₂ S (0-1,000ppm) ≤35 seconds		conds	
	СО	<30 seconds	H ₂ S (0-5,000ppm	H ₂ S (0-5,000ppm) ≤40 seconds		conds	
			H ₂ S (0-10,000pp	H ₂ S (0-10,000ppm) ≤40		40 seconds	
Cell lifetime	O ₂ cell is 3 years in	air, all other cells 2 years in	air				

^{*}Plus accuracy of calibration gas used

© Product designs and specifications are subject to change without notice. User is responsible for determining suitability of product.

^{**}Times are taken from the point gas enters the BIOMETHANE 3000 module. Sample times will vary depending on length of sample pipe

¹ The process will be paused during an auto calibration

☐ BIOMETHANE 3000

TECHNICAL SPECIFICATIONS CONTINUED

PUMP				
Flow	300ml / minute typically. Please note that the default operation of the pump is always off and uses the positive pressure of the gas at the sample point			
Flow-fail point	Flow rate less than 75ml / minute or vacuum greater than 350 mbar			
Maximum vacuum restart	-375 mbar			
COMMUNICATIONS				
Output channels	Up to six analogue 4-20mA output channels that are user configurable for current sink or source inputs plus Modbus RTU over RS-485			
	Optional Profibus, Profinet or Ethernet module			
Alarm notifications	1 x fault relay			
	7 x user-configurable alarms that can trigger a relay when above or below a set value and one to inform the operator of the results of the autocalibration. In addition, one can be used to indicate to the operator when the catchpot is full and requires emptying			
Relay outputs	Single pole changeover 6A 24Vdc relay volt free			
ENVIRONMENT CONDIT	TIONS			
Operating pressures	-350 mbar to +350 mbar*			
IP rating	IP65			
Humidity	0-95% non-condensing humidity			
PHYSICAL				
Size	650 x 600 x 210mm (with supplied wall mounting brackets) per enclosure (2 enclosures)			
Weight	Maximum 36.5kg per enclosure			
Enclosure	Stainless steel, 600 x 600 x 210mm, IP65 rated			
Operation keys	Alpha-numeric keypad with 'tactile' membrane			
Display	480 x 272 pixel RGB TFT display, 96mm x 55mm			
Moisture removal filters	User replaceable microfibre filter and 2.0µm PTFE water traps			
Heater	100W mains powered ATEX certified heater for 110V or 230V mains supply			
CERTIFICATION RATING				
ISO17025	Calibrated under UKAS accreditation (certificate number 4533)			
ATEX / IECEx marking				
BS EN 61010-1:2010	Safety requirements for electrical equipment for measurement, control, and laboratory use			
BS EN 50270:2006	Electromagnetic compatibility- electrical apparatus for the detection and measurement of combustible gases, toxic gases or oxygen			

^{*}Pressures will need regulating in order not to damage the system. This is the responsibility of the user.











